

APPARATUS

| CODE | PART NAME | REMARKS |
|------------|---------------|--|
| ATG-00-013 | atg-8l | Tag Closed Cup Flash Point Tester, Metal Bath Model, External Chiller excluded |
| ATG-00-014 | atg-8w | Tag Closed Cup Flash Point Tester, Water Bath Model |

STANDARD ACCESSORIES

| CODE | PART NAME | Qty | REMARKS |
|------------|--|-----|----------------------------------|
| ATG-02-411 | Test Cup for Tag | 1 | |
| 210-00-061 | O-Ring, G-45 | 1 | For Test Cup |
| 320-00-035 | Thermofuse | 1 | Pack of 5, with Insulating Tube. |
| ---- | Gas Hose $\phi 9 \times \phi 16$, 1.5m, with Hose Bands | 1 | |
| ABL-01-401 | Windscreen for abl-8/atg-8 | 1 | |
| ---- | Braided Hose with Wire bands, 1.5m | 1 | For atg-8w only |
| ---- | Drain Hose with Hose Band | 1 | For atg-8w only |
| ATG-04-021 | Flow Control Valve | 1 | For atg-8w only |
| 210-02-191 | Insulated Tubing, 2m | 2 | For atg-8l only |
| 170-01-051 | Hose Band, SPN-8 | 4 | For atg-8l only |
| MPC-02-032 | AC Power Cord | 1 | 3.0m, For 100 to 120V |
| MPC-02-033 | | | 2.5m, For 220 to 240V |
| ---- | atg-8w/l Manuals | 1 | Manuals for atg-8w and atg-8l. |

OPTIONAL ACCESSORIES

| CODE | PART NAME | REMARKS |
|------------|-------------------|---|
| 070-00-068 | Printer, BS2-80TS | Prints out test data and instrument settings. |
| 070-00-281 | Mini USB Keyboard | Sample ID entering |
| | CHILLER | For atg-8l only, Temperature range +5 to 55°C |

SUGGESTED SPARES for 2 YEARS

| CODE | PART NAME | Qty | CODE | PART NAME | Qty |
|------------|------------------|-----|------------|--------------------------------|-----|
| ATG-02-411 | Test Cup for Tag | 1 | ABL-01-003 | Electric Igniter, EL-8 | 2 |
| 210-00-061 | O-Ring, G-45 | 1 | ATG-04-031 | Temperature Sensor for atg-8 | 1 |
| 320-00-035 | Thermofuse | 1 | ABL-01-012 | Flash Detector for abl-8/atg-8 | 1 |

We reserve the right of changes without prior notice. Printed in Japan 1201(E)

DISTRIBUTED BY


TANAKA SCIENTIFIC
 7-10-3 Ayase, Adachi-ku, Tokyo 120-0005 JAPAN
 TEL: +81-3-3620-1711 FAX: +81-3-3620-1713
 URL: <http://www.tanaka-sci.com>
 E-mail: overseas-group@tanaka-sci.com

TANAKA Automatic Petroleum Tester

AUTOMATED TAG CLOSED CUP FLASH POINT TESTER

atg-8w/atg-8l



All New

Model atg-8 series is the automated Tag Closed Cup flash point tester from Tanaka. Based on 45 years experience of manufacturing the automated flash point testers, innovative 8th generation series has been released. The theme of 8th generation is "Ergonomic, Versatile and Premium".

- ERGONOMIC COLOR-COORDINATED DISPLAY FOR HIGH VISIBILITY FOR EVERYBODY (UNIVERSAL DESIGN) AND INTUITIVE SWITCH OPERATION.
- ENHANCED VERSATILITY INCLUDING USB PORT FOR OPTIONAL FLASH MEMORY OR KEYBOARD, PASSWORD PROTECTION AND DATA STRAGE.
- ALL-IN-ONE CHASSIS DESIGN: COMPACT, LIGHT-WEIGHT AND PREMIUM EXTERIOR DESIGN.
- CONFORMS TO STANDARD TEST METHOD (ASTM D56, IP304)

SINGLE ACTION SETTING

The cup cover is permanently mounted on a swing-arm assembly, which allows an easy handling of the test cup cover. It makes free from handling the hot cup cover after completing a test.

The swing arm assembly can be lifted up to vertical position for easier test cup cleaning.

EASY OPERATION

Select a test mode and enter the expected flash point; while the instrument executes the test, you are free to do other lab work. The instrument follows the exact procedures prescribed in the test method, and the completion of the test cycle is signaled by beeps. The test result is brightly shown on the color LCD module.

INTERCHANGEABLE IGNITION SOURCE

atg-8w/atg-8l is equipped with both gas and electric igniter as the ignition source. It only takes a few minutes to switch from gas to electric or vice versa.

DRAFT PROTECTION SHIELD COVER

A FM4910 conforming cover insures the flame integrity. This material shall meet the numerical criterion of each index. (1)FPI(Fire Propagation Index) ≤ 6 (2)SDI(Smoke Damage Index) ≤ 0.4

DATA HANDLING

The unit is equipped to store up to 200 successive test results and export to an optional printer or to a LIMS through an RS-232 port or USB memory.

USB PORT

A USB port allows connecting USB keyboard for easy entering sample name and connecting USB flash memory for data copy to a PC or test conditions/parameters copy to the tester.



COMPACT DESIGN & ENERGY EFFICIENT

atg-8l uses Peltier Cells for sample cooling/heating. Just a small chiller with water suffices the cooling requirement. No methanol is required.

A control part and a test part are put into one body which culminated in achieved in the light weighting of the unit. A liquid-cooling metal bath model and conventional water bath model are available by different measurement temperature range.

SPECIFICATIONS

| | |
|---|---|
| TYPE | All –in one, microprocessor controlled |
| TEST STANDARDS | ASTM D56, IP304 |
| MEASURING RANGE | atg-8w: Ambient to +95°C atg-8l: -20°C to +70°C at cooling medium of 10°C 20°C to +95°C at cooling medium of 50°C |
| MEASUREMENT MODES | ASTM CRM, ASTM Normal, ASTM SPE (search), ASTM Rapid and User custom |
| BATH | atg-8w:Water atg-8l: Metal Block |
| SAMPLE COOLING/HEATING | atg-8w:by Plate Heater/by by replacing Bath Liquid atg-8l: by Peltier Devices |
| DISPLAY | 5.7 inch (117 x 88mm) color LCD with LED back light |
| TEMPERATURE SENSOR | PT-100 in stainless steel sheath |
| FLASH DETECTOR | CRC Thermocouple |
| IGNITION SOURCE | Gas ignition with automatic lighting or electric ignition. Interchangeable |
| BAROMETRIC CORRECTION | Automatic correction by built-in barometric pressure sensor |
| I/O PORT | RS-232C = 1 channel (for PC or Optional Printer) USB=1 channel (for USB keyboard or USB Flash Memory) |
| SAFETY SHUTDOWN | Automatically shuts off and the problem is reported by buzzer and display, in case: (a) EFP+10°C or 95°C is reached, (b) temperature sensor is found defective, (c) flash detector is found defective, (d) thermofuse is blown, (e) electric Igniter is blown, (f) built-in battery is found drained out, (g) test cover is not set in place, (h)control computer runs away (no display) |
| GAS SUPPLY (When Gas Ignition is used) | LP gas or natural gas (Max. pressure:10kPa) |
| POWER SUPPLY | 100/120VAC or 220/240 VAC, 50/60Hz |
| POWER CONSUMPTION | atg-8w: 500VA MAX. atg-8l: 250VA MAX. |
| ELECTRICITY CONSUMPTION | atg-8w: 40Wh for 1 test (about 30 min), 15.12gCO ₂ (@ 0.378kgCO ₂ /1kWh) atg-8l: 30Wh for 1 test (about 30 min), 11.34gCO ₂ (@ 0.378kgCO ₂ /1kWh) |
| AMBIENT TEMP. RANGE | 10°C to 35°C |
| DIMENSIONS & WEIGHT | 230mm (W) × 470mm (D) × 390mm (H) 17kg |

CHILLER REQUIREMENTS for atg-8l

| | |
|--------------------------|---|
| TYPE | Circulator with heating/cooling capacities. Open bath type. |
| TEMPERATURE RANGE | 5°C to +55°C or wider |
| PUMP CAPACITY | Pressure:0.2bar minimum Flow rate:2L/min minimum |
| COOLING CAPACITY | 250W @10°C minimum |
| HEATER CAPACITY | 500W minimum |
| BATH VOLUME | 3L minimum |

Julabo F25-ED is suitable for this specification.